# Systematic Review and Meta-Analysis of Antigen Detection Tests for the Diagnosis of Tuberculosis<sup>▽</sup>†

L. L. Flores, <sup>1</sup>‡ K. R. Steingart, <sup>2</sup>‡\* N. Dendukuri, <sup>3</sup> I. Schiller, <sup>3</sup> J. Minion, <sup>3</sup> M. Pai, <sup>3,4</sup> A. Ramsay, <sup>5</sup> M. Henry, <sup>6</sup> and S. Laal <sup>7,8,9</sup>

Division of Pulmonary and Critical Care Medicine, San Francisco General Hospital, University of California, San Francisco, California<sup>1</sup>;

Department of Health Services, University of Washington School of Public Health, Seattle, Washington<sup>2</sup>; Department of Epidemiology,

Biostatistics & Occupational Health, McGill University, Montreal, Quebec, Canada<sup>3</sup>; Respiratory Epidemiology and

Clinical Research Unit, Montreal Chest Institute, Montreal, Canada<sup>4</sup>; UNICEF/UNDP/World Bank/WHO

Special Programme for Research and Training in Tropical Diseases (TDR), World Health Organization,

Geneva, Switzerland<sup>5</sup>; Department of Community Health Systems, School of Nursing, University of

California, San Francisco, California<sup>6</sup>; Departments of Pathology<sup>7</sup> and Microbiology,<sup>8</sup>

New York University Langone Medical Center, New York, New York; and

Veterans Affairs Medical Center, New York, New York<sup>9</sup>

Received 2 June 2011/Returned for modification 12 July 2011/Accepted 2 August 2011

Tests that detect Mycobacterium tuberculosis antigens in clinical specimens could provide rapid direct evidence of active disease. We performed a systematic review to assess the diagnostic accuracy of antigen detection tests for active tuberculosis (TB) according to standard methods and summarized test performance using bivariate random effects meta-analysis. Overall, study quality was a concern. For pulmonary TB (47 studies, 5,036 participants), sensitivity estimates ranged from 2% to 100% and specificity from 33% to 100%. Lipoarabinomannan (LAM) was the antigen most frequently targeted (23 studies, 49%). The pooled sensitivity of urine LAM was higher in HIV-infected than HIV-uninfected individuals (47%; 95% confidence interval [CI], 26 to 68% versus 14%; 95% CI, 4 to 38%); pooled specificity estimates were similar: 96%; 95% CI, 81 to 100% and 97%; 95% CI, 86 to 100%, respectively. For extrapulmonary TB (21 studies, 1,616 participants), sensitivity estimates ranged from 0% to 100% and specificity estimates from 62% to 100%. Five studies targeting LAM, ESAT-6, Ag85 complex, and the 65-kDa antigen in cerebrospinal fluid, when pooled, yielded the highest sensitivity (87%; 95% CI, 61 to 98%), but low specificity (84%; 95% CI, 60 to 95%). Because of the limited number of studies targeting any specific antigen other than LAM, we could not draw firm conclusions about the overall clinical usefulness of these tests. Further studies are warranted to determine the value of LAM detection for TB meningitis in high-HIV-prevalence settings. Considering that antigen detection tests could be translated into rapid point-of-care tests, research to improve their performance is urgently needed.

The World Health Organization (WHO) estimates that in 2009, 9.4 million new cases of tuberculosis (TB) occurred and 1.7 million people died of the disease (89). The vast majority of these patients live in low- and middle-income countries (LMIC) where TB diagnosis depends primarily on smear microscopy. Microscopy has low sensitivity and does not detect smear-negative TB (77), which may account for 24% to 61% of all pulmonary cases in HIV-infected individuals (27). Improved diagnostic tests, such as mycobacterial culture and nucleic acid amplification (NAA) tests, are available in highincome countries but are often too expensive and complex for routine use by TB control programs in resource-constrained settings in which TB is endemic. Lack of access to diagnostic services in resource-limited settings presents an additional barrier to using these tests. The Xpert MTB/RIF (Cepheid, Inc., Sunnyvale, CA), recently endorsed by the WHO, is rapid and

highly sensitive for detection of TB and drug resistance; however, this new technology is costly, preventing its use in many areas where the epidemic is most severe (10). Accurate, rapid, inexpensive, and simple diagnostic tests are urgently needed for TB care and control.

There have been considerable efforts over the past 50 years to devise a rapid TB test based on antibody detection. However, substantial published evidence suggests that current commercial serological tests are inaccurate for diagnosis of active TB (22, 73–76, 91). A meta-analysis of noncommercial serological tests identified several potential candidate antigens and antigen combinations for inclusion in an antibody detection test for pulmonary TB (PTB) (73). Recently, serological profiling of the complete *Mycobacterium tuberculosis* proteome has been described (38). In July 2011, the WHO issued a policy statement against the use of commercial serological tests for the diagnosis of active pulmonary and extrapulmonary TB (EPTB) while stressing the importance of continued research on antibody-based detection tests and point-of-care tests (90).

Another approach aims to detect circulating mycobacterial antigens in clinical specimens such as serum, sputum, urine, cerebrospinal spinal fluid (CSF), and pleural fluid. A common design for antigen detection tests is the sandwich enzymelinked immunosorbent assay (ELISA) technique, also called

<sup>\*</sup> Corresponding author. Mailing address: Department of Health Services, University of Washington School of Public Health, Seattle, WA 98195-7230. Phone: (646) 243-9043. Fax: (775) 254-3124. E-mail: karenst@uw.edu.

<sup>‡</sup> These authors contributed equally to this work.

<sup>†</sup>Supplemental material for this article may be found at http://cvi.asm.org/.

<sup>&</sup>lt;sup>▽</sup> Published ahead of print on 10 August 2011.




FIG. 1. Antigen-capture ELISA. (Far left) Antigen-specific monoclonal or polyclonal antibodies are coated onto the surface of plastic wells. (Second from left) The specimen containing the antigen to be measured is added. The antigen released is then "captured" by the antigen-specific antibodies. (Third from left) The captured antigen(s) are detected by secondary antibodies using the same antibody or an antibody directed toward an epitope that is different from the epitope recognized by the capture antibody. These secondary antibodies can be directly labeled with moieties, like biotin, or enzymes, like horseradish peroxidase. When biotin is used, visualization requires the addition of streptavidin conjugated to enzyme (horseradish peroxidase or alkaline phosphatase). (Far right) The appropriate substrate is added. The result is visualized by the color generated and measured with an ELISA plate reader. (Reprinted from reference 29a with permission of the publisher.)

the antigen-capture ELISA, described in Fig. 1. Another format is the lateral-flow immunochromatographic assay, sometimes referred to as a dipstick. In one common scheme of this technique, the antigen is initially introduced onto the device by adding a sample from a clinical specimen, such as urine or serum. Free antigen, if present, binds to an antibody/microsphere (bead) complex, providing a visually detectable color change (44). Dipsticks require little and sometimes no sample processing. Agglutination is an alternate method that can detect antigens when they clump together (agglutinate) with antibodies and form a visual complex.

In comparison with conventional diagnostics, antigen detection tests appear to offer several advantages. Immunochromatographic tests, for example, are rapid (results may be available within minutes) and easy to operate. If developed into a point-of-care test, an antigen detection test could extend TB diagnosis to remote health posts. Antigen detection tests provide direct evidence of active disease, thus allowing for immediate initiation of TB treatment. An antigen detection test using a specimen such as urine would be particularly attractive with children, who may have difficulty providing sputum. Finally, in patients suspected of extrapulmonary TB, an antigen detection test might prevent the use of more invasive tests.

To our knowledge, two systematic reviews have previously assessed the diagnostic accuracy of antigen detection tests for TB (22, 45). Both reviews reported on the performance of tests targeting lipoarabinomannan (LAM), a major glycolipid component of the cell wall of *M. tuberculosis* and other mycobacteria (12, 30). The first review identified one study evaluating sputum LAM; the sensitivity was 63%, and the specificity 92% (22). The second review focused on urine LAM (9 studies) (45). Depending on the classification of culture-negative clinical TB, pooled sensitivity estimates ranged from 34% to 60% and pooled specificity from 93% to 94%. Two recent narrative reviews were also identified (16, 55). However, one review focused only on urine-based tests (55), and neither review appraised study quality or performed a meta-analysis.

The objective of the current systematic review is to estimate the diagnostic accuracy of antigen detection tests using different clinical specimens for PTB and EPTB in adults and children with and without HIV infection.

### MATERIALS AND METHODS

We followed methods for conducting and reporting systematic reviews and meta-analyses recommended by the Cochrane Collaboration Diagnostic Test Accuracy Working Group and the PRISMA statement, including the preparation of a protocol (26, 40, 46).

Selection criteria and definitions. (i) Types of studies. Included were diagnostic studies (with any study design) that evaluated antigen detection tests (using immunologic methods such as a sandwich ELISA) for PTB and EPTB in patients providing clinical specimens prior to or within 14 days of starting antituberculous treatment.

- (ii) Participants. Participants included adults and children with suspected or confirmed active TB (at least 10 cases), with or without HIV infection.
- (iii) Intervention (test). The tests consisted of an antigen detection test using an immunologic method, such as a sandwich ELISA or immunochromatographic assav.
- (iv) Target conditions. Target conditions consisted of pulmonary TB and extrapulmonary TB.
- (v) Reference standards. Positivity on culture was used for PTB and positivity on culture, smear, or histopathological examination for EPTB. For TB meningitis, we accepted positivity by commercial NAA tests because these tests provide high specificity (22, 51).
- (vi) Outcomes. Sensitivity refers to the proportion of patients among TB patients with a positive antigen detection test result confirmed by the reference standard. Specificity refers to the proportion of non-TB participants with a negative antigen detection test result among non-TB participants. To estimate specificity, we selected only one non-TB group if a study had more than one such group. The preferred non-TB participants were those in whom active TB was initially suspected but later ruled out ("other respiratory disease" or "other disease" groups) and who were from the same population as TB patients.

EPTB was classified as lymph node, pleural, meningeal, and/or central nervous system, bone and/or joint, genitourinary, abdominal, skin, other site, disseminated, and multiple (extrapulmonary TB cases from different sites are combined to obtain at least 10 extrapulmonary TB cases).

Countries were classified as high income or LMIC according to the World Bank list of economies (88).

The following were excluded: (i) studies published before 1990, (ii) animal studies, (iii) conference abstracts and proceedings, (iv) studies of the detection of latent TB infection, (v) studies of the detection of nontuberculous mycobacteria, and (vi) basic science literature focused on the detection/cloning of new antigens or evaluation of their immunological properties (i.e., early preclinical studies).

Search methods. Database searching was performed (PubMed, 1 January 1990 to 16 April 2011, and Biosis, Embase, and Web of Science, 1 January 1990 to 10 February 2010) for relevant studies that reported data on immunological tests (based either on antigen or antibody detection) for active TB. For studies prior to 2006, databases were searched only in English because of limited resources, and for those after 2006, in all languages (see supplemental material S1). Reference lists of included papers and reviews were hand searched, and investigators contacted to identify additional studies.

Study selection. Initially, two reviewers (L.L.F. and K.R.S.) independently screened citations for relevance and then, based upon prespecified selection




FIG. 2. Flow of studies in the review of antigen detection tests for the diagnosis of pulmonary and extrapulmonary TB. \*, two of the papers in the extrapulmonary TB group are also included in the pulmonary TB group.

criteria, independently reviewed full-text articles. Disagreements about study selection were resolved by discussion between the reviewers.

**Data extraction.** A data extraction form was created, pilot tested with a subset of eligible studies, and then finalized. Two reviewers (L.L.F. and K.R.S.) independently extracted data from the included studies with the standardized form on the following characteristics: study design, age group (children < 15 years), HIV status, country, smear status, form of TB, assay type, test name, antigen identity, type of specimen, sensitivity, and specificity. When possible, studies were classified by HIV or smear status. Differences between reviewers concerned mainly methodological quality and were resolved by discussion. When necessary, authors were contacted for additional information.

We looked for information on patient-important outcomes, such as mortality, time to diagnosis, and number of patient visits during the diagnostic process.

Assessment of methodological quality. Two reviewers independently assessed study quality using QUADAS (Quality Assessment of Diagnostic Accuracy Studies), a validated tool to evaluate the presence of bias in diagnostic accuracy studies (86) (see supplemental material S2).

**Data analysis.** Descriptive analyses were performed using SPSS (version 14.0.1.366). For each study, the sensitivity and specificity of the test along with 95% confidence intervals (CI) were calculated, and forest plots created to display the estimates using Review Manager 5.1 (the Nordic Cochrane Center, Copenhagen, Denmark). Heterogeneity was assessed by visual examination of forest plots.

Selection of subgroups for meta-analysis. Heterogeneity was addressed by prespecifying subgroups by type of specimen evaluated, antigen(s) detected, and smear and HIV status (see Fig. S1 in the supplemental material). For the meta-analysis, at least four studies for inclusion in a subgroup were required.

Bivariate meta-analyses that jointly modeled both sensitivity and specificity were performed. These models weighted studies according to the sampling variability within studies as well as the unexplained heterogeneity between studies using a random effects approach (63). The model was estimated using a Bayesian approach with nonsubjective prior distributions and implemented using WinBUGS (version 1.4.1) (71). The meta-analysis provides a pooled estimate of

the sensitivity and specificity across studies. A hierarchical summary receiver operating characteristic (HSROC) plot provides added information when there is high heterogeneity between studies, and the pooled estimate may not be representative of all studies. The HSROC curve plots sensitivity versus specificity and provides information on the overall performance of a test across different thresholds. The closer the curve is to the upper left hand corner of the plot (sensitivity and specificity are both 100%), the better is the performance of the test. The plots were made using R (version 2.6.1) (58).

# **RESULTS**

Initially, 8,263 citations were identified (Fig. 2). After screening titles and abstracts, 132 potentially relevant full-text articles were retrieved; 49 publications describing a total of 68 studies (47 for PTB and 21 for EPTB) were selected for inclusion in the review. A list of excluded articles with reasons for exclusion is available upon request from the authors.

**PTB.** (i) Characteristics of included studies. Thirty papers, 29 in English (1, 2, 5, 6, 8, 9, 14, 15, 17–20, 23, 33, 37, 39, 43, 49, 50, 54, 57, 60, 61, 64, 66, 67, 70, 72, 85) (2 papers by Shah et al. concerned the same investigation) and one in French (24), were included (Table 1). The 47 PTB studies involved 5,036 participants (sample size = 6,540). The majority (81%) of the studies were conducted in LMIC. Eight studies involved HIV-infected individuals; none involved children. In each study, all non-TB participants were from the same country as TB patients. Twenty-three (49%) studies had a cross-sectional design, and 21 (45%) studies a case-control design. In three

TABLE 1. Characteristics of included studies evaluating antigen detection tests for the diagnosis of pulmonary tuberculosis<sup>d</sup>

Specificity (%) (95% CI)	96 (78–100) 89 (79–96) 80 (66–90)	90 (73–98)	77 (56–91)	77 (56–91)	82 (72–90)			97 (92–99)	95 (91–98)	95 (91–98)	70 (61–79) 70 (61–79)		80 (65–90)	93 (88–97)		96 (78–100) 92 (79–98) 89 (82–93)	89 (82–93)	7 (2–20) 16 (9–24) 86 (76–93) 100 (84–100)
Sensitivity (%) (95% CI)	28 (18–41) 91 (86–95) 96 (79–100)	(20–22)	94 (73–100)	47 (24–71)	100 (86–100)	57 (34–78)	79 (63–90)	48 (32–63)	83 (64–94)	60 (26–88)	93 (81–99) 83 (63–95)	100 (86–100)	94 (82–99)	97 (90–100)	95 (88–99) 98 (93–100)	22 (13–34) 63 (41–81) 89 (84–93)	82 (74–88)	100 (92–100) 84 (74–91) 96 (91–99) 94 (73–100)
No. of patients with/without TB	64/23 175/65 24/49	41/30	18/26	19/26	24/74 50/63	21/63	30/37 42/201	44/120	29/176	10/176	43/111 24/111	$24/45^a$	47/45 <sup>a</sup>	$69/164^a$	79/21 79/21 98/102	64/23 24/38 198/131	116/131	44/41 80/109 110/78 18/21
Test name/antigen(s) detected (H37Rv designation)	In-house/45/47 kDa (1860) In-house/20 kDa (not found) In-house/Ag85 complex	(3804c and 1886c) In-house/M. tuberculosis	antigens, unspecified Latex Agglutination Test for TB/M. tuberculosis	antigens, unspecified Latex Agglutination Test for TB/M. tuberculosis	antigens, unspecified In-house/65 kDa (0440) In-house/LAM	In-house/LAM	In-nouse/FLD In-house/M. tuberculosis	antigens, unspecified In-house/38 kDa (0934)	In-house/M. tuberculosis	antigens, unspecified In-house/M. tuberculosis	anugens, unspecmed Patho-TB/proprietary Patho-TB/proprietary	In-house/M. tuberculosis	antigens, unspecified In-house/M. tuberculosis	In-house/PPD	Pathouse/ATD Pathou TB/proprietary Diagnos TB Ag/LAM, 30 FDa (1886c) and 2	proprietary antigens In-house/45/47 kDa (1860) In-house/LAM In-house/19 (3763), 38	(0934), and 58 kDa (0440) In-house/19 (3763), 38	(1954), and 58 KDa (1944) Clearview TB ELISA/LAM Clearview TB ELISA/LAM Patho-TB/proprietary In-house/LAM
s Smear status Comparison group	ORD ORD OD	ORD	ORD	ORD	OD ORD	ORD	nealuiy OD	OD and healthy	ORD	ORD	ORD	ORD	ORD	ORD	ORD ORD	ORD ORD ORD	ORD	ORD ORD ORD OD
Smear status	Smear+ Unspecified Smear+	Unspecified	Smear+	Smear-	Smear+	Unspecified	Smear+	Unspecified	Smear+	Smear-	Smear+ Smear-	Smear+	Smear-	Smear+	Smear+ Unspecified	Smear+ Unspecified Smear+	Smear-	Unspecified Unspecified Unspecified Unspecified
HIV statu	HIV- Not reported HIV-	Not reported	Not reported	Not reported	Not reported HIV—	HIV+	Not reported	Not reported	Not reported	Not reported	HIV- HIV-	Not reported	Not reported	Not reported	Not reported HIV+	HIV – Not reported Not reported	Not reported	HIV+ HIV- Not reported Not reported
Country	Madagascar Egypt India	Russia	Bangladesh	Bangladesh	India Mexico	Mexico	Romania	The Netherlands	Saudi Arabia	Saudi Arabia	Iran Iran	Kuwait	Kuwait	Thailand	Tunisia India	Madagascar Korea India	India	South Africa South Africa France Vietnam
Author (reference) (year, study) <sup>e</sup>	Chanteau (15) (2000, b) El-Masry (23) (2008) Kashyap (33) (2007)	Khomenko (37) (1996)	Mamun (43) (1990, a)	Mamun (43) (1990, b)	Rajan (57) (2007) Sada (64) (1992 a)	Sada (64) (1992, b)	Stavri (72) (1990)	Verbon (85) (1993)	al-Orainey (2) (1992, a)	al-Orainey (2) (1992, b)	Alavi-Naini (1) (2009, a) Alavi-Naini (1) (2009, b)	Araj (5) (1993, a)	Araj (5) (1993, b)	Banchuin (6) (1990, a)	Ben-Selma (8) (2009) Chakraborty (14) (2009)	Chanteau (15) (2000, a) Cho (17) (1997) Deodhar (19) (1998, a)	Deodhar (19) (1998, b)	Dheda (20) (2010, c) Dheda (20) (2010, d) Fabre (24) (2007) <sup>b</sup> Pereira Arias-Bouda
Specimen type (no. of studies)	Serum (12)								Sputum (20)									

88 (82–92) 88 (82–92) 100 (91–100) 99 (95–100) 100 (98–100) 89 (81–94) 89 (81–94) 100 (79–100) 91 (79–98) 84 (68–94) 84 (68–94) 96 (91–99) 33 (24-44) 33 (24-44) 33 (21-46) 50 (41-60) 28 (15-44) 38 (15-65) 24 (7-50) 74 (55-88) 42 (20-67) 64 (52-75) 53 (42-64) 76 (62–87) 22 (9–42) 52 (30–47) 20 (10–35) 6 (2–14) 65 (49–78) 50 (31-69) 82/82 50/82 27/155 21/155 21/155 44/41 88/109 88/104 40/114 40/114 16/16 11/45 31/37 73/122 88/122 48/90 28/90 Clearview TB ELISA/LAM Clearview TB ELISA/LAM Clearview TB ELISA/LAM Clearview TB ELISA/LAM 30 Diagnos TB Ag/LAM, 30 proprietary antigens Diagnos TB Ag/LAM, proprietary antigens kDa (1886c) and 2 kDa (1886c) and 2 LAM- ELISA/LAM LAM- ELISA/LAM LAM-ELISA/LAM LAM-ELISA/LAM LAM-ELISA/LAM LAM-ELISA/LAM LAM-ELISA/LAM In-house/(1656) LAM-ELISA/LAM LAM-ELISA/LAM ORD ORD Unspecified Unspecified Unspecified Smear+ Unspecified Unspecified Unspecified Smear+ mear+ Smear+ Smear+ Smear-Smear-Smear-Not reported HIV + + H HIV + H HIV + H HIV + H HIV + HIV+ HIV. Zimbabwe US and Brazil Tanzania India South Africa South Africa South Africa South Africa South Africa Zimbabwe Fanzania l'anzania Tanzania Tanzania anzania Reither (60) (2009, a) Reither (60) (2009, b) Reither (60) (2009, c) Mutetwa (49) (2009, a) Mutetwa (49) (2009, b) Daley (18) (2009, a)
Daley (18) (2009, b)
Dheda (20) (2010, a)
Dheda (20) (2010, b) Reither (61) (2010, a) Napolitano (50) (2008) Reither (61) (2010, b) Shah (66) (2010, a) Shah (66) (2010, b)  $awn (39) (2009)^c$ Urine (15)

<sup>c</sup> Specimen was unconcentrated urine.
<sup>d</sup> ELISA, enzyme-linked immunosorbent assay; LAM, lipoarabinomannan; OD, other diseases; ORD, other respiratory diseases; TB, tuberculosis; PPD, purified protein derivative; HIV±, people with and without HIV infection.
<sup>e</sup> Lowercase letters following the year of study indicate distinct studies, as some papers may have included more than one study. Cases included both treated and untreated patients; results were nearly equivalent when stratified by treatment status. a Unit of analysis is specimen rather than patient

TABLE 2. Antigen detection tests for the diagnosis of pulmonary TB

	•	
Name(s) of antigen(s) detected (H37Rv designation for protein antigens)	No. of studies	Reference(s)
Lipoarabinomannan	20	9, 17, 18, 20, 39, 49, 54, 60, 64, 66
Mycobacterium tuberculosis antigens, unspecified	8	2, 5, 37, 43, 72
Proprietary	4	1, 8, 24
Lipoarabinomannan, 30 kDa (1886), and 2 other proprietary	3	14, 61
Purified protein derivative (PPD)	3	6, 70
45/47 kĎa (1860)	2	15
19 (3763), 38 (0934), and 58 kDa (0440)	2 2	19
20 kDa (Rv not found)	1	23
Ag85 complex (3804c and 1886c)	1	33
Rv1656	1	50
38 kDa (0934)	1	85
65 kDa (0440)	1	57

studies, study design was unclear. The median number of TB patients included in each study was 43 (interquartile range of 24 to 71).

Of the total 47 studies, 36 (77%) used ELISA, 7 (15%) an immunochromatographic test, and 4 an agglutination test. As shown in Table 2, single antigens were targeted in 24 (51%) studies and multiple antigens in 21 studies; the number of antigens was not reported in two studies. Of the 47 studies, LAM was the antigen most frequently targeted, either as a single antigen or one of several antigens (23 studies, 49%). Among the 20 studies targeting LAM alone, urine was the specimen most often evaluated (14 studies). In 15 (32%) studies, antigen identity was not reported. Five commercial tests were evaluated, accounting for 25 (53%) of the total studies: Patho-TB (Anda Biologicals, France; 4 studies), Diagnos TB Ag (Biomed Industries, India; 3 studies), LAM-ELISA (Chemogen, United States; 10 studies [this assay was a precommercial prototype and the test is no longer available]), Clearview TB ELISA (Inverness Medical Innovations, United States; 6 studies), and a latex agglutination test for TB (Wellcome Diagnostics, United Kingdom; 2 studies).

As assessed with QUADAS, all studies satisfied six quality items: acceptable reference standard (a criterion for inclusion in the review), acceptable delay between tests, partial verification avoided, incorporation avoided, reference standard results blinded (as the culture result was considered to be entirely objective in interpretation), and relevant clinical information. However, studies had serious shortcomings in two essential items, representative spectrum (selection of a representative patient population) and blinding, factors that have been shown empirically to be associated with exaggerated accuracy estimates (42): only 18 (38%) studies involved a representative patient population and only 19 (40%) reported blinding of the result of the antigen detection test (see Fig. S2 in the supplemental material).

Figure 3 is a forest plot of the sensitivity and specificity of




FIG. 3. Forest plots of sensitivity and specificity for antigen detection tests for pulmonary TB, all studies. For letters preceding author names: A, serum; B, sputum; C, urine. (Lowercase letters following year of study indicate distinct studies.) TP, true positive; FP, false positive; FN, false negative; TN, true negative. The 95% confidence intervals (CI) are included between the brackets. The figure shows the sensitivity and specificity estimates for individual studies (squares) and 95% CIs (black horizontal lines).

individual studies stratified by specimen type. There was considerable heterogeneity within strata (for example, within the serum subgroup, sensitivity ranged from 28% to 100%). Estimated sensitivity ranged from 2% to 100%, and specificity from 33% to 100% across studies. Similar ranges of performance (sensitivity of 6% to 100% and specificity of 33% to 100%) were seen when the analysis was restricted to studies from LMIC (data not shown).

(ii) Urine LAM studies of HIV-infected and uninfected TB patients. As shown in Fig. 4, five studies reported the performance of urine LAM in individuals with TB-HIV coinfection; four of these studies reported data comparing LAM antigen detection in individuals with and without HIV infection. All four studies found improved sensitivity in HIV-infected com-

pared with HIV-uninfected individuals (14% to 53% higher sensitivity), although in three studies, specificity was lower by 6% to 7%, and in one study, specificity was increased by 1%. In three studies involving HIV-TB-coinfected patients, results were stratified by CD4 cell count (20, 39, 67). All three studies found a higher sensitivity of urine LAM in patients with lower CD4 cell counts. We identified only one study that evaluated urine LAM in sputum smear-negative HIV-infected individuals. Sensitivity was 42%, with a 95% CI of 20 to 67%, and specificity was 84%, with a 95% CI of 68 to 94% (60).

Lawn et al. found that only 2 h were required for detection of LAM in urine compared with an average of 24 days for detection of *M. tuberculosis* complex by culture (39). In addition, five of 22 patients at risk for paradoxical TB immune




FIG. 4. Forest plots of sensitivity and specificity for urine LAM for pulmonary TB. Studies of HIV-infected patients are identified with "\*." TP, true positive; FP, false positive; FN, false negative; TN, true negative. 95% confidence intervals (CI) are included between the brackets. The figure shows the sensitivity and specificity estimates for individual studies (squares) and 95% CIs (black horizontal lines).

reconstitution disease had detectable LAM. The authors noted a trend toward increased mortality in patients with LAM positivity, although this trend was not statistically significant (39).

(iii) Meta-analysis of studies of PTB. Among the 47 studies of PTB patients, there was considerable heterogeneity in specimen type, antigen identity, and smear and HIV status. We were able to identify five subgroups that were more homogeneous (Tables 3 and 4). For sputum LAM (studies of both smear-positive and -negative patients), the pooled sensitivity for LAM was 87%, with a 95% CI of 71 to 96%, and the pooled specificity 70%, with a 95% CI of 6 to 99% (Table 3). For urine LAM, the performances were similar in smear-positive (pooled sensitivity, 54%, and 95% CI, 18 to 86%; pooled specificity, 90%, and 95% CI, 83 to 95%) and smear-negative (pooled sensitivity, 51%, and 95% CI, 18 to 83%; pooled specificity, 90%, and 95% CI, 79 to 96%) patients (Table 4). Within studies of HIV-uninfected TB patients (n = 4), the pooled sensitivity of urine LAM was 14%, with a 95% CI of 4 to 38%, and the pooled specificity was 97%, with a 95% CI of 86 to 100%. In HIV-infected TB patients, the pooled sensitivity was higher but still low (n = 5, 47%; 95% CI, 26 to 68%), and pooled specificity similar (96%; 95% CI, 81 to 100%) (Table 5). The estimated probability that the pooled sensitivity of the test with HIV-infected patients exceeds that of the test with HIV-uninfected patients was 98%.

**EPTB.** (i) Characteristics of included studies. All 19 articles were written in English (4, 6, 13, 21, 29, 32, 34, 35, 47, 48, 52, 53, 56, 66, 68, 69, 82, 84, 87) (Table 5). The 21 EPTB studies involved 1,616 participants (sample size = 1,690). The majority (91%) of studies were conducted in LMIC. In four studies, HIV-infected individuals comprised 49% (21), 68% (52), 84%

(53), and 85% (67) of the total enrolled patients. The remaining papers did not report HIV status. No studies involved children. In each study, all non-TB participants were from the same country as TB patients. The median number of TB patients included in each study was 22 (interquartile range of 14 to 31). Seventeen (81%) studies involved in-house tests, and four used the Clearview TB ELISA. Of the total 21 studies, 14 (64%) used the ELISA, 2 (9%) an agglutination test, and 5 immunohistochemical staining of biopsy specimens. In seven (33%) studies, the targeted antigen was unspecified.

Overall, studies had serious limitations: only three (14%) studies were considered to involve a representative patient population and only nine (43%) reported blinding of the antigen detection test result, (see Fig. S3 in the supplemental material). Figure 5 is a forest plot of the sensitivity and specificity of individual studies stratified by specimen type. There was considerable heterogeneity within strata (for example, within the biopsy specimen subgroup, sensitivity ranged from 36% to 100%). The estimated sensitivity for all studies ranged from 0% to 100%, and specificity from 62% to 100%. In the EPTB group, no studies reported patient-important outcomes.

(ii) Meta-analysis of studies of EPTB. Among the 21 studies of EPTB patients, there was considerable heterogeneity in disease site, specimen type, and antigen identity. We were able to identify three subgroups that were homogenous in terms of specimen type: serum, CSF, and biopsy specimen (Table 3). Studies targeting LAM, ESAT-6, Ag85 complex, and the 65-kDa antigen (65-kDa) in CSF, when pooled, yielded the highest sensitivity (87%; 95% CI, 61 to 98%), but low specificity (84%; 95% CI, 60 to 95%) (Table 3; Fig. 6).

TABLE 3. Pooled sensitivity and specificity estimates by subgroup, with posterior medians

Subgroup or specimen	No. of studies	No. of participants	Sensitivity (%) (95% CI)	Specificity (%) (95% CI)
Pulmonary tuberculosis Sputum lipoarabinomannan (LAM)	4	461	87 (71–96)	70 (6–99)
Extrapulmonary tuberculosis				
Serum <sup>a</sup>	4	246	76 (24–97)	89 (74–96)
Cerebrospinal fluid <sup>b</sup>	5	404	87 (61–98)	84 (60–95)
Biopsy/immunohistochemical staining <sup>c</sup>	4	174	80 (23–99)	86 (69–97)

 $<sup>^{</sup>a}$  Antigens detected: excretory secretory (ES)-20 (n = 1); ES-31 (n = 20); EST-6 [contains 38 kDa (Rv0934) and 41 kDa], ES-20, and ES-31 (n = 1); the precise identities of ES20, ES-31, and EST6 are not known.

<sup>&</sup>lt;sup>b</sup> Antigens detected: LAM (n = 2); Ag85complex (Rv3804c and Rv1886c, n = 1); ESAT-6 early secreted antigenic target-6 (Rv3875, n = 1); 65 kDa (Rv0440, n = 1). Pooled accuracy estimates are based on 5 studies with defined antigen targets. When data from an additional study targeting "nonspecific glycolipid antigens" were included (82), pooled sensitivity was 90% (56 to 99%) and pooled specificity 86% (72 to 94%).

<sup>&</sup>lt;sup>c</sup> Antigens were unspecified in all studies.

TABLE 4. Pooled sensitivity and specificity estimates for urine lipoarabinomannan (LAM) detection<sup>a</sup>

Pulmonary TB subgroup	No. of studies	No. of participants	Sensitivity (%) (95% CI)	Specificity (%) (95% CI)
Smear-positive	6	906	54 (18–86)	90 (83–95)
Smear-negative	5	728	51 (18–83)	90 (79–96)
HIV-infected	5	844	47 (26–68)	96 (81–100)
HIV-uninfected	4	357	14 (4–38)	97 (86–100)

<sup>&</sup>lt;sup>a</sup> Posterior medians with 95% credible intervals are given in parentheses.

## DISCUSSION

Forty-seven PTB and 21 EPTB studies evaluating antigen detection tests for the diagnosis of TB were included in this systematic review. Eight subgroups comprising 37 studies were selected for meta-analysis. Within these subgroups, for both PTB and EPTB, sensitivity and specificity were variable. At a recent meeting convened by Médecins Sans Frontières, Partners in Health, and the Treatment Action Group, a multidisciplinary group of TB experts, including clinicians, test developers, laboratory experts, research scientists, and community representatives, proposed a set of minimal specifications for a point-of-care test based on modeling and expert opinion: in adults with pulmonary TB, the test should achieve a sensitivity of 95% in smear-positive and 60% to 80% in smear-negative cases and a specificity of 95%; and in extrapulmonary TB, the test should provide a sensitivity of 80% and specificity of 95% (41). In this review, for both PTB and EPTB, pooled accuracy estimates in all analyses fell short of achieving the required sensitivity and specificity values. However, we caution that our meta-analyses are based on a small number of studies, meaning that there is still considerable uncertainty in the usefulness or lack of usefulness of these tests. A new TB test may have several purposes, for example, triage, replacement test for microscopy, or add-on test following microscopy. Although we




FIG. 6. Summary HROC plot of sensitivity and specificity for antigen detection tests of cerebrospinal fluid for the diagnosis of tuberculous meningitis. The width of the circles is proportional to the number of patients in each study. The square is the summary value for sensitivity and specificity.




FIG. 5. Forest plots of sensitivity and specificity for antigen detection tests for the diagnosis of extrapulmonary TB, all studies. For letters preceding author names: A, biopsy specimen; B, cerebrospinal fluid; C, lymph node aspirate; D, pleural fluid; E, serum; F, urine. TP, true positive; FP, false positive; FN, false negative; TN, true negative. The 95% confidence intervals (CI) are included between the brackets. The figure shows the sensitivity and specificity estimates for individual studies (squares) and 95% CIs (black horizontal lines).

TABLE 5. Characteristics of included studies evaluating antigen detection tests for the diagnosis of extrapulmonary tuberculosis<sup>a</sup>

1624

Specimen type	Author (reference) (year, study group)	Form of tuberculosis	Country	HIV status	Comparison group	Test name/antigen(s) detected (H37Rv designation)	No. of patients with/without TB	Sensitivity (%) (95% CI)	Specificity (%) (95% CI)
Biopsy	Carabias (13) (1998)	Multiple and	Spain	Not reported	QO	In-house/M. tuberculosis antioens, unspecified	11/38	36 (11–69)	76 (60–89)
	Mukherjee (48) (2002)	Lymph node	India	Not reported	OD	In-house/M. tuberculosis	50/10	74 (60–85)	100 (69–100)
	Radhakrishnan (56)	Meningeal	India	Not reported	OD	In-house/M. tuberculosis	10/14	100 (69–100)	100 (77–100)
	(1991) Wiley (87) (1990)	Not reported	USA	Not reported	ОО	anugens, unspectned In-house/ <i>M. tuberculosis</i> antigens, unspecified	13/28	92 (64–100)	82 (63–94)
Cerebral spinal fluid	Kashyap (34) (2007)	Meningeal	India	Not reported	OD	In-house/Ag85 complex	27/87	89 (71–98)	90 (81–95)
	Kashyap (35) (2009) Mudaliar (47) (2006)	Meningeal Meningeal	India India South Africa	Not reported Not reported	000	(3804¢ and 10805) In-house/ESAT-6 (3875) In-house/65 kDa (0440)	15/65 18/72	100 (78–100) 100 (81–100)	94 (85–98) 89 (79–95)
	Fatel (32) (2009) Patel (53) (2010) Tiwari (82) (2007)	Meningeal Meningeal Meningeal	South Africa India	HIV± Not reported	388	Clearview TB ELISA/LAM Clearview TB ELISA/LAM In-house/glycolipid cocktail	39/54 302/79	04 (33–87) 69 (52–83) 99 (97–100)	02 (32–80) 65 (51–77) 94 (86–98)
Lymph node aspirate	Goel (29) (2008) Jain (32) (2003)	Lymph node Lymph node	India India	Not reported Not reported	00 00	In-house/38 kDa (0934) In-house/38 kDa (0934) and crude antigens	48/23 34/62	100 (93–100) 100 (90–100)	100 (85–100) 90 (80–96)
Pleural fluid	Anie (4) (2007)	Pleural	India	Not reported	ORD	In-house/tuberculosis- associated glycolipid	11/71	100 (72–100)	100 (95–100)
	Banchuin (6) (1990, c) Dheda (21) (2009)	Pleural Pleural	Thailand South Africa	Not reported $HIV\pm$	ORD	protein In-house/PPD Clearview TB ELISA/LAM	26/11 12/10	12 (2–30) 8 (0–38)	100 (72–100) 100 (69–100)
Serum	Shende (68) (2007, a) Shende (68) (2007, b) Shende (69) (2008) Upadhye (84) (2007)	Lymph node Lymph node Lymph node Multiple*	India India India India	Not reported Not reported Not reported Not reported	0D 0D 0RD 0D	In-house/ES-31 In-house/ES-20 In-house/ES-20 In-house/ES-31, ES-43, EST-6	22/15 22/15 30/35 32/75	73 (50–89) 82 (60–95) 90 (73–98) 47 (29–65)	87 (60–98) 87 (60–98) 89 (73–97) 91 (82–96)
Urine	Banchuin (6) (1990, d) Shah (66) (2010, c)	Pleural Disseminated	Thailand South Africa	Not reported HIV±	ORD OD	In-house/PPD Clearview TB ELISA/LAM	26/11 18/122	0 (0–13) 78 (52–94)	100 (72–100) 96 (91–99)
Juil Commence A 21 12 a	Od moon the decomment bester omme	doso co mente de care de	itachi ononati	I ton moniton to wi	DOAT FORT 6	our famories of autism not basum. DCAT & Dr. 2075, souly consisted autismuis transit & DCT & coutsing 39 (D-4002A) and 41 LDn. I AM	TOT 19 town	20 (Dri0024) o	2d 41 LDo: I AM

"ELISA, enzyme-linked immunosorbent assay; ES, excretory [precise identity of antigen not known]; ESAT-6 (Rv3875), early secreted antigenic target-6; EST-6 contains 38 (Rv0934) and 41 kDa: LAM, lipoarabinomannan; OD, other diseases; ORD, other respiratory diseases; TB, tuberculosis; PPD, purified protein derivative; HIV±, people with and without HIV infection; \*, includes lymph node (n = 8), meningeal (n = 5), bone and joint (n = 6), abdominal (n = 5), pleural (n = 4), and disseminated (n = 4) TB.

1625

did not identify any studies that specifically considered the added value of antigen detection tests in addition to microscopy, in such a scenario, the required sensitivity for smearpositive patients may not need to be that great or necessarily higher than that for smear-negative patients.

LAM was the most frequently studied antigen, probably because it is a major constituent of the mycobacterial cell wall (12, 30), and a commercial LAM test has been developed. Although in this review the sensitivity of LAM detection was higher in sputum (87%) than urine (53%), urine LAM has important implications. First, it provides proof of principle that antigen detection with an easily obtained specimen can be exploited to develop a TB diagnostic test. Second, our metaanalysis demonstrates that urine LAM provides higher sensitivity in HIV-infected (47%) compared with HIV-uninfected (14%) TB patients, as was also found in a recent systematic review of urine LAM for TB (45). One possible explanation may be that the former set of patients has a higher or more disseminated bacterial burden with LAM antigenemia, leading to detectable LAM in urine. Of note, LAM positivity appeared to be inversely related to the CD4 count. Third, recent studies evaluating NAA tests with multiple specimens from the same patients have shown that this strategy increases the sensitivity of bacterial detection (59), suggesting that evaluation of both sputum and urine from the same patient may provide higher sensitivity than either specimen alone. Fourth, although in HIV-infected individuals we found the sensitivity of urine LAM for pulmonary TB to be relatively low, the combination of LAM testing and smear microscopy may provide additional benefit as proposed by Shah and colleagues (66). Although pooled specificity of urine LAM was relatively high ( $\geq 96\%$ ), specificity was found to be low (84% to 94%) in several studies and even lower in studies of serum and sputum LAM. Possible reasons include the presence of latent TB infection, misclassification of subclinical TB cases, or cross-reactivity with nontuberculous mycobacteria (3, 28, 30, 36). Studies are now required to clarify the cause of reduced LAM specificity.

For EPTB, the highest pooled sensitivity was achieved in tests for TB meningitis, a condition that is challenging to diagnose. Microscopy has poor sensitivity (~5%) for TB meningitis, and culture takes several weeks for results (80). Commercial NAA tests show high specificity (98%) and low and variable sensitivity (approximately 60%), require technical skill, and are expensive (51). With reported fatality rates above 50% (81), even a modest improvement in the accurate diagnosis of TB meningitis would be valuable. Two studies evaluating CSF LAM were well designed and reported (52, 53). Both involved a high proportion (>60%) of HIV-infected TB patients and achieved sensitivities of approximately 65%, suggesting that CSF LAM might be useful as a rule-in test for the rapid diagnosis of TB meningitis in high-HIV-prevalence settings. A point-of-care version of a LAM detection test for TB meningitis might be good to explore, especially if the test combined detection of LAM and other promising biomarkers, such as gamma interferon and adenosine deaminase (22,

Antigen detection tests for TB will require identification of antigens that have the following properties: abundant expression by in vivo bacteria, presence of antigens in the extracellular environment, and resistance to degradation by the enzymes that may be released during the inflammatory responses of the host. Moreover, antigens that are specific to M. tuberculosis will likely be necessary. Considering that M. tuberculosis adapts to its environment by modifying its transcriptional responses (25, 65, 78), it is not unreasonable to assume that the optimal antigens or antigen combinations for diagnosing different forms of EPTB may vary. Antigen detection in serum may be more difficult due to complex formation with the antibodies they elicit. The feasibility of using antigen detection tests for paucibacillary TB is unclear, but tests that detect multiple antigens may yield higher sensitivities than tests that detect single antigens. Besides LAM, no other antigens with potential value were identified in this review.

The strengths of our review include the use of standard guidelines, a comprehensive search strategy, two independent reviewers for all stages of the review process, and quality assessment with QUADAS. Although heterogeneity between diagnostic accuracy studies is to be expected (31), we attempted to address heterogeneity by prespecifying similar subgroups and by using a bivariate meta-analysis method.

Our review had limitations. The meta-analysis was limited by the small number of studies of any particular antigen detection test, with the exception of LAM. Our review did not account for different specimen processing techniques that may have contributed to variability in test performance. For example, 16 (80%) of the total 20 studies evaluating sputum used a processing method (studies differed with respect to the chemicals used and duration and gravitational force/speed of centrifugation), one study did not use a processing method, and three studies did not report whether or not a processing method was used. In studies evaluating urine LAM, two tests, the LAM-ELISA and the Clearview TB ELISA, were used. Although these assays contain the same polycolonal antibodies, the manufacturing processes differ; factors such as pH and viscosity may have affected diagnostic accuracy. In a systematic review of urine LAM for active TB, there was no statistically significant difference in performance noted between the groups of studies using fresh and previously frozen urine (45). In about 30% of the studies, test performance was reported, but antigen identity was unspecified. We were limited in our assessment of study quality by incomplete reporting of information. We encourage investigators to be more transparent about their studies. Publication bias is possible because studies with poor performance were unlikely to be unpublished. Statistical tests, such as funnel plots, are not recommended to detect potential publication bias in meta-analyses of diagnostic data (79). Accuracy estimates in our meta-analysis may be exaggerated because of the risk of bias from the selection of unrepresentative samples and lack of blinding (42, 62). The small sample size of individual studies contributes to heterogeneity between them. Going forward, researchers of primary diagnostic studies should follow published guidelines for conducting and reporting their work to ensure that their efforts contribute to a high-quality evidence base (7, 11).

In conclusion, because of the limited number of studies targeting any specific antigen other than LAM and concerns about methodological quality in a majority of studies, we could not draw firm conclusions about the clinical usefulness of antigen detection tests. Further studies are warranted to deter-

mine the value of LAM detection for TB meningitis in high-HIV-prevalence settings. Future studies are also needed that evaluate the performance of antigen detection tests in HIV-infected individuals with smear-negative TB. Considering that antigen detection tests could be translated into rapid and inexpensive point-of-care tests, research to improve their performance is urgently needed.

### ACKNOWLEDGMENTS

This systematic review was commissioned and funded by USAID through the UNICEF/UNDP/World Bank/WHO Special Programme for Research and Training in Tropical Diseases (TDR).

We thank Dick Menzies, McGill University, for his helpful comments on the manuscript.

Conflict of interests: K.R.S. serves as Coordinator of the Evidence Synthesis and Policy Subgroup of Stop TB Partnership's New Diagnostics Working Group (NDWG). A.R. served as the secretary of the NDWG (2006 to 2010) and is employed by WHO/TDR, the agency that administered the grant used to fund this systematic review. A.R. contributed to the conception and design of the systematic review and critical revision of and decision to publish the manuscript. J.M. is a recipient of the Quebec Respiratory Health Training Fellowship. M.P. is a recipient of a New Investigator Award from the Canadian Institutes of Health Research (CIHR). This funding source had no role in the preparation of the manuscript nor the decision to submit the manuscript for publication. M.P. serves as an external consultant for the Bill & Melinda Gates Foundation. M.P. also serves as a cochair of the NDWG. We have no financial involvement with any organization or entity with a financial interest in or financial conflict with the subject matter or materials discussed in the manuscript apart from those disclosed.

#### REFERENCES

- Alavi-Naini, R., M. Metanat, E. Alijani, and H. Mozaffar. 2009. Patho-TB test for the rapid diagnosis of pulmonary tuberculosis. J. Res. Med. Sci. 14:301–307
- al-Orainey, I. O., M. O. Gad el Rab, M. S. al-Hajjaj, and E. S. Saeed. 1992.
   Detection of mycobacterial antigens in sputum by an enzyme immunoassay.
   Eur. J. Clin. Microbiol. Infect. Dis. 11:58–61.
- Alvarez-Uria, G., et al. 2009. Non-tuberculous mycobacteria in the sputum of HIV-infected patients: infection or colonization? Int. J. STD AIDS 20:193– 105
- Anie, Y., et al. 2007. Diagnostic approaches in patients with tuberculous pleural effusion. Diagn. Microbiol. Infect. Dis. 59:389–394.
- Araj, G. F., B. H. Fahmawi, T. D. Chugh, and M. Abu-Salim. 1993. Improved detection of mycobacterial antigens in clinical specimens by combined enzyme-linked immunosorbent assay. Diagn. Microbiol. Infect. Dis. 17:119– 127.
- Banchuin, N., S. Wongwajana, U. Pumprueg, and J. Jearanaisilavong. 1990.
   Value of an ELISA for mycobacterial antigen detection as a routine diagnostic test of pulmonary tuberculosis. Asian Pac. J. Allergy Immunol. 8:5–11.
- Banoo, S., et al. 2006. Evaluation of diagnostic tests for infectious diseases: general principles. Nat. Rev. Microbiol. 4:S20–S32.
- Ben-Selma, W., et al. 2009. Rapid detection of Mycobacterium tuberculosis in sputum by Patho-TB kit in comparison with direct microscopy and culture. Diagn. Microbiol. Infect. Dis. 65:232–235.
- Boehme, C., et al. 2005. Detection of mycobacterial lipoarabinomannan with an antigen-capture ELISA in unprocessed urine of Tanzanian patients with suspected tuberculosis. Trans. R. Soc. Trop. Med. Hyg. 99:893–900.
- Boehme, C. C., et al. 2010. Rapid molecular detection of tuberculosis and rifampin resistance. N. Engl. J. Med. 363:1005–1015.
   Bossuyt, P. M., et al. 2003. Towards complete and accurate reporting of
- Bossuyt, P. M., et al. 2003. Towards complete and accurate reporting of studies of diagnostic accuracy: the STARD Initiative. Ann. Intern. Med. 138:40–44.
- Brennan, P. J. 2003. Structure, function, and biogenesis of the cell wall of Mycobacterium tuberculosis. Tuberculosis (Edinb.) 83:91–97.
- Carabias, E., E. Palenque, R. Serrano, J. M. Aguado, and C. Ballestin. 1998. Evaluation of an immunohistochemical test with polyclonal antibodies raised against mycobacteria used in formalin-fixed tissue compared with mycobacterial specific culture. APMIS 106:385–388.
- Chakraborty, N., et al. 2009. A rapid immunochromatographic assay for the detection of Mycobacterium tuberculosis antigens in pulmonary samples from HIV seropositive patients and its comparison with conventional methods. J. Microbiol. Methods 76:12–17.
- 15. Chanteau, S., et al. 2000. 45/47 kilodalton (APA) antigen capture and anti-

- body detection assays for the diagnosis of tuberculosis. Int. J. Tuberc. Lung Dis. 4:377–383.
- Cho, S. N. 2007. Current issues on molecular and immunological diagnosis of tuberculosis. Yonsei Med. J. 48:347–359.
- Cho, S. N., et al. 1997. Detection of Mycobacterium tuberculosis antigens in sputum for the diagnosis of pulmonary tuberculosis. J. Korean Soc. Microbiol. 32:285–291.
- Daley, P., et al. 2009. Blinded evaluation of commercial urinary lipoarabinomannan for active tuberculosis: a pilot study. Int. J. Tuberc. Lung Dis. 13:989–995.
- Deodhar, L., A. Gogate, R. C. Padhi, and C. R. Desai. 1998. Standardization of a dot blot immunoassay for antigen detection in cases of pulmonary tuberculosis & its evaluation with respect to the conventional techniques. Indian J. Med. Res. 108:75–79.
- Dheda, K., et al. 2010. Clinical utility of a commercial LAM-ELISA assay for TB diagnosis in HIV-infected patients using urine and sputum samples. PLoS One 5:e9848.
- Dheda, K., et al. 2009. Clinical diagnostic utility of IP-10 and LAM antigen levels for the diagnosis of tuberculous pleural effusions in a high burden setting. PLoS One 4:e4689.
- Dinnes, J., et al. 2007. A systematic review of rapid diagnostic tests for the detection of tuberculosis infection. Health Technol. Assess. 11:1–196.
- El-Masry, S., I. El-Kady, M. H. Zaghloul, and M. K. Al-Badrawey. 2008. Rapid and simple detection of a mycobacterium circulating antigen in serum of pulmonary tuberculosis patients by using a monoclonal antibody and Fast-Dot-ELISA. Clin. Biochem. 41:145–151.
- Fabre, M., et al. 2007. [Assessment of the Patho-TB kit for diagnosis of tuberculosis]. Pathol. Biol. (Paris) 55:482–485. (In French.)
- Fontan, P., V. Aris, S. Ghanny, P. Soteropoulos, and I. Smith. 2008. Global transcriptional profile of Mycobacterium tuberculosis during THP-1 human macrophage infection. Infect. Immun. 76:717–725.
- Gatsonis, C., and P. Paliwal. 2006. Meta-analysis of diagnostic and screening test accuracy evaluations: methodologic primer. Am. J. Roentgenol. 187: 271–281.
- Getahun, H., M. Harrington, R. O'Brien, and P. Nunn. 2007. Diagnosis of smear-negative pulmonary tuberculosis in people with HIV infection or AIDS in resource-constrained settings: informing urgent policy changes. Lancet 369:2042–2049.
- Glatman-Freedman, A., J. M. Martin, P. F. Riska, B. R. Bloom, and A. Casadevall. 1996. Monoclonal antibodies to surface antigens of Mycobacterium tuberculosis and their use in a modified enzyme-linked immunosorbent spot assay for detection of mycobacteria. J. Clin. Microbiol. 34:2795–2802.
- Goel, M. M., and P. Budhwar. 2008. Species-specific immunocytochemical localization of Mycobacterium tuberculosis complex in fine needle aspirates of tuberculous lymphadenitis using antibody to 38 kDa immunodominant protein antigen. Acta Cytol. 52:424–433.
- 29a. Goldsby, R. A., T. J. Kindt, B. A. Osborne, and J. Kuby. 2003. Immunology, 5th ed. W. H. Freeman and Company, New York, NY.
- Hamasur, B., G. Kallenius, and S. B. Svenson. 1999. A new rapid and simple method for large-scale purification of mycobacterial lipoarabinomannan. FEMS Immunol. Med. Microbiol. 24:11–17.
- Harbord, R. M., J. J. Deeks, M. Egger, P. Whiting, and J. A. Sterne. 2007. A unification of models for meta-analysis of diagnostic accuracy studies. Biostatistics 8:239–251.
- Jain, A., R. K. Verma, V. Tiwari, and M. M. Goel. 2003. Development of a new antigen detection dot-ELISA for diagnosis of tubercular lymphadenitis in fine needle aspirates. J. Microbiol. Methods 53:107–112.
- Kashyap, R. S., et al. 2007. Diagnosis of tuberculosis in an Indian population by an indirect ELISA protocol based on detection of antigen 85 complex: a prospective cohort study. BMC Infect. Dis. 7:74.
- Kashyap, R. S., et al. 2007. Comparison of an adenosine deaminase assay with ELISA for the diagnosis of tuberculous meningitis infection. Med. Sci. Monit. 13:BR200–BR204.
- Kashyap, R. S., et al. 2009. Diagnostic value of early secreted antigenic target-6 for the diagnosis of tuberculous meningitis patients. Infection 37: 500 512
- Khatter, S., U. B. Singh, J. Arora, T. Rana, and P. Seth. 2008. Mycobacterial infections in human immuno-deficiency virus seropositive patients: role of non-tuberculous mycobacteria. Indian J. Tuberc. 55:28–33.
- Khomenko, A. G., et al. 1996. Serodiagnosis of tuberculosis: detection of mycobacterial antibodies and antigens. Tuber. Lung Dis. 77:510–515.
- Kunnath-Velayudhan, S., et al. 2010. Dynamic antibody responses to the Mycobacterium tuberculosis proteome. Proc. Natl. Acad. Sci. U. S. A. 107: 14703–14708.
- Lawn, S. D., et al. 2009. Urine lipoarabinomannan assay for tuberculosis screening before antiretroviral therapy diagnostic yield and association with immune reconstitution disease. AIDS 23:1875–1880. doi:10.1097/ QAD.0b013e32832e05c8.
- Leeflang, M. M., J. J. Deeks, C. Gatsonis, and P. M. Bossuyt. 2008. Systematic reviews of diagnostic test accuracy. Ann. Intern. Med. 149:889–897.
- Lemaire, J. F., and M. Casenghi. 2010. New diagnostics for tuberculosis: fulfilling patient needs first. J. Int. AIDS Soc. 13:40.

- Lijmer, J. G., et al. 1999. Empirical evidence of design-related bias in studies of diagnostic tests. JAMA 282:1061–1066.
- Mamun, K. Z., and P. Shears. 1990. Investigation of a latex slide agglutination technique in the diagnosis of pulmonary tuberculosis, in Dhaka, Bangladesh. Trop. Doct. 20:79–81.
- McNerney, R., and P. Daley. 2011. Towards a point-of-care test for active tuberculosis: obstacles and opportunities. Nat. Rev. Microbiol. 9:204–213.
- Minion, J., et al. 4 July 2011. Diagnosing tuberculosis with urine lipoarabinomannan: systematic review and meta-analysis. Eur. Respir. J. [Epub ahead of print.] doi:10.1183/09031936.00025711.
- Moher, D., A. Liberati, J. Tetzlaff, and D. G. Altman. 2009. Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. PLoS Med. 6:e1000097.
- Mudaliar, A. V., R. S. Kashyap, H. J. Purohit, G. M. Taori, and H. F. Daginawala. 2006. Detection of 65 kD heat shock protein in cerebrospinal fluid of tuberculous meningitis patients. BMC Neurol. 6:34.
- Mukherjee, A., N. Kalra, and N. R. Beena. 2002. Immunohistochemical detection of mycobacterial antigen in tuberculous lymphadenitis. Indian J. Tuberc. 49:213–216.
- Mutetwa, R., et al. 2009. Diagnostic accuracy of commercial urinary lipoarabinomannan detection in African tuberculosis suspects and patients. Int. J. Tuberc. Lung Dis. 13:1253–1259.
- Napolitano, D. R., N. Pollock, S. S. Kashino, V. Rodrigues, Jr., and A. Campos-Neto. 2008. Identification of Mycobacterium tuberculosis ornithine carboamyltransferase in urine as a possible molecular marker of active pulmonary tuberculosis. Clin. Vaccine Immunol. 15:638–643.
- Pai, M., et al. 2003. Diagnostic accuracy of nucleic acid amplification tests for tuberculous meningitis: a systematic review and meta-analysis. Lancet Infect. Dis. 3:633–643.
- Patel, V. B., et al. 2009. Utility of a novel lipoarabinomannan assay for the diagnosis of tuberculous meningitis in a resource-poor high-HIV prevalence setting. Cerebrospinal Fluid Res. 6:13.
- 53. Patel, V. B., et al. 2010. Comparison of a clinical prediction rule and a LAM antigen-detection assay for the rapid diagnosis of TBM in a high HIV prevalence setting. PLoS One 5:e15664.
- Pereira Arias-Bouda, L., et al. 2000. Development of antigen detection assay for diagnosis of tuberculosis using sputum samples. J. Clin. Microbiol. 38: 2278–2283.
- Peter, J., et al. 2010. Urine for the diagnosis of tuberculosis: current approaches, clinical applicability, and new developments. Curr. Opin. Pulm. Med. 16:262–270.
- Radhakrishnan, V. V., A. Mathai, N. S. Radhakrishnan, D. Rout, and S. Sehgal. 1991. Immunohistochemical demonstration of mycobacterial antigens in intracranial tuberculoma. Indian J. Exp. Biol. 29:641–644.
- Rajan, A. N., R. S. Kashyap, H. J. Purohit, G. M. Taori, and H. F. Daginawala. 2007. Serodiagnosis of tuberculosis based on the analysis of the 65 kD heat shock protein of Mycobacterium tuberculosis. Int. J. Tuberc. Lung Dis. 11:792–797.
- 58. R Development Core Team. 2010. A language and environment for statistical computing. R Foundation for Statistical Computing, Vienna, Austria. http://www.R-project.org. Accessed 17 September 2010.
   59. Rebollo, M. J., et al. 2006. Blood and urine samples as useful sources for the
- Rebollo, M. J., et al. 2006. Blood and urine samples as useful sources for the direct detection of tuberculosis by polymerase chain reaction. Diagn. Microbiol. Infect. Dis. 56:141–146.
- Reither, K., et al. 2009. Low sensitivity of a urine LAM-ELISA in the diagnosis of pulmonary tuberculosis. BMC Infect. Dis. 9:141.
- Reither, K., et al. 2010. Evaluation of Diagnos TB AG, a flow-through immunoassay for rapid detection of pulmonary tuberculosis. Int. J. Tuberc. Lung Dis. 14:238–240.
- Rutjes, A. W., et al. 2006. Evidence of bias and variation in diagnostic accuracy studies. CMAJ 174:469–476.
- Rutter, C. M., and C. A. Gatsonis. 2001. A hierarchical regression approach to meta-analysis of diagnostic test accuracy evaluations. Stat. Med. 20:2865– 2884.
- Sada, E., D. Aguilar, M. Torres, and T. Herrera. 1992. Detection of lipoarabinomannan as a diagnostic test for tuberculosis. J. Clin. Microbiol. 30:2415– 2419.
- Schnappinger, D., et al. 2003. Transcriptional adaptation of Mycobacterium tuberculosis within macrophages: insights into the phagosomal environment. J. Exp. Med. 198:693–704.
- Shah, M., et al. 2010. Quantitative analysis of a urine-based assay for detection of lipoarabinomannan in patients with tuberculosis. J. Clin. Microbiol. 48:2972–2974.
- Shah, M., et al. 2009. Diagnostic accuracy of a urine lipoarabinomannan test for tuberculosis in hospitalized patients in a high HIV prevalence setting. J. Acquir. Immune Defic. Syndr. 52:145–151.

- 68. Shende, N., V. Upadhye, S. Kumar, N. Gangane, and B. C. Harinath. 2007. Study of M. tuberculosis ES-31 and ES-20 antigen levels in different pathogenic grades of lymph node tuberculosis. Int. J. Tuberc. Lung Dis. 11:222–226
- Shende, N., V. Upadhye, S. Kumar, and B. C. Harinath. 2008. A low molecular weight ES-20 protein released in vivo and in vitro with diagnostic potential in lymph node tuberculosis. Indian J. Med. Microbiol. 26:29–33.
- Sood, J., et al. 1991. Penicillinase ELISA for detection of tubercular antigen in tuberculosis. J. Commun. Dis. 23:173–177.
- Spiegelhalter, D., A. Thomas, N. Best, and D. Lunn. 2004. WinBUGS user manual, version 1.4.1. http://www.mrc-bsu.cam.ac.uk/bugs. Accessed 13 May 2010
- Stavri, D., et al. 1990. Specific antibodies and mycobacterial antigens in patient sera with pulmonary tuberculous and nontuberculous diseases. Note II. Arch. Roum. Pathol. Exp. Microbiol. 49:331–338.
- Steingart, K. R., et al. 2009. Performance of purified antigens for serodiagnosis of pulmonary tuberculosis: a meta-analysis. Clin. Vaccine Immunol. 16:260–276.
- Steingart, K. R., et al. 2011. Commercial serological tests for the diagnosis of active pulmonary and extrapulmonary tuberculosis: an updated systematic review and meta-analysis. PLoS Med. 8(8):e1001062.
- Steingart, K. R., et al. 2007. Commercial serological antibody detection tests for the diagnosis of pulmonary tuberculosis: a systematic review. PLoS Med. 4:e202.
- Steingart, K. R., et al. 2007. A systematic review of commercial serological antibody detection tests for the diagnosis of extrapulmonary tuberculosis. Thorax 62:911–918.
- Steingart, K. R., et al. 2006. Sputum processing methods to improve the sensitivity of smear microscopy for tuberculosis: a systematic review. Lancet Infect. Dis. 6:664–674.
- Talaat, A. M., R. Lyons, S. T. Howard, and S. A. Johnston. 2004. The temporal expression profile of Mycobacterium tuberculosis infection in mice. Proc. Natl. Acad. Sci. U. S. A. 101:4602–4607.
- Tatsioni, A., et al. 2005. Challenges in systematic reviews of diagnostic technologies. Ann. Intern. Med. 142:1048–1055.
- Thwaites, G. E., et al. 2002. Diagnosis of adult tuberculous meningitis by use of clinical and laboratory features. Lancet 360:1287–1292.
- Thwaites, G. E., et al. 2004. Dexamethasone for the treatment of tuberculous meningitis in adolescents and adults. N. Engl. J. Med. 351:1741–1751.
- Tiwari, R. P., S. K. Garg, R. N. Bharmal, S. Kartikeyan, and P. S. Bisen. 2007. Rapid liposomal agglutination card test for the detection of antigens in patients with active tuberculosis. Int. J. Tuberc. Lung Dis. 11:1143–1151.
- Tuon, F. F., et al. 2010. Adenosine deaminase and tuberculous meningitis—a systematic review with meta-analysis. Scand. J. Infect. Dis. 42:198–207.
- 84. Úpadhye, V., N. Shende, S. Kumar, and B. C. Harinath. 2007. Detection of antibody and antigen in extrapulmonary tuberculosis patients' sera using a cocktail of mycobacterial excretory secretory antigens and their antibodies. Biomed. Res. 18:161–166.
- Verbon, A., et al. 1993. Evaluation of different tests for the serodiagnosis of tuberculosis and the use of likelihood ratios in serology. Am. Rev. Respir. Dis. 148:378–384.
- 86. Whiting, P., A. W. Rutjes, J. B. Reitsma, P. M. Bossuyt, and J. Kleijnen. 2003. The development of QUADAS: a tool for the quality assessment of studies of diagnostic accuracy included in systematic reviews. BMC Med. Res. Methodol. 3:25.
- 87. Wiley, E. L., T. J. Mulhollan, B. Beck, J. A. Tyndall, and R. G. Freeman. 1990. Polyclonal antibodies raised against Bacillus Calmette-Guerin, Mycobacterium duvalii, and Mycobacterium paratuberculosis used to detect mycobacteria in tissue with the use of immunohistochemical techniques. Am. J. Clin. Pathol. 94:307–312.
- World Bank. 2010. World bank list of economies. World Bank, Washington, DC. http://siteresources.worldbank.org/DATASTATISTICS/Resources/CLASS .XLS. Accessed 5 August 2010.
- World Health Organization. 2010. Global tuberculosis control: WHO report 2010 WHO/HTM/TB/2010.7. WHO, Geneva, Switzerland.
- World Health Organization. 2011. Policy statement: commercial serodiagnostic tests for diagnosis of tuberculosis. WHO/HTM/TB/2011.5.
   WHO, Geneva, Switzerland. http://whqlibdoc.who.int/publications/2011/9789241502054\_eng.pdf. Accessed 29 July 2011.
- World Health Organization on behalf of the Special Programme for Research and Training in Tropical Diseases. 2008. Laboratory-based evaluation of 19 commercially available rapid diagnostic tests for tuberculosis. WHO, Geneva, Switzerland.
- Xu, H. B., R. H. Jiang, L. Li, W. Sha, and H. P. Xiao. 2010. Diagnostic value of adenosine deaminase in cerebrospinal fluid for tuberculous meningitis: a meta-analysis. Int. J. Tuberc. Lung Dis. 14:1382–1387.